



PRODUCT CERTIFICATION GUIDE

Ohio Semitronics, Inc. offers five types of certification covering the calibration of products on new orders. To order one of the options below, include the document number (shown in bold at the left) on your purchase order. Consult OSI Sales for pricing as it varies with the type of cert and the product complexity. Special data-points are available.

A-7003-01 Certificate of Compliance (C of C). Standard with all orders/OSi products at no charge.

Includes: Company name, P.O. #, Date, Model(s), Serial number(s).

Document states: It is hereby certified that the above stated models, in the quantities listed, are in full compliance with all applicable requirements and specifications.

Configuration, operation and safety characteristics, of the stated models, have been tested and inspected to verify compliance with published specifications and any additional requirements as specified by the referenced purchase order.

Accuracy of factory test and measurement equipment is established through regular comparisons traceable to recognized national or international measurement standards such as those maintained by the National Institute of Standards and Technology (N.I.S.T.)

Signature: Quality Assurance personnel, Group supervisor, or Dept. Mgr.

Note: For model MCCT metering class current transformers listing compliance with IEEE/ANSI C57.13-2008, use cert **A-7003-191**

A-7003-02 C of C with Documented Traceability.

Includes: All the information and statements listed on above document and traceability to N.I.S.T. provided via equipment and report identification numbers listed on certificate. 1 document per unit/serial number.

Signature: Q.A. Manager.

A-7003-03 C of C with Traceability and Data Points.

Includes: All the information and statements listed on the above document and a 10 point calibration data recorded after the final calibration.

The format and price of the data can vary with the type of equipment and/or customer requirements.

Signature: Quality Assurance Manager.

A-7003-04 C of C with ISO 17025 Certificate of Calibration.

Includes: Contracted Calibration service provider to provide ISO/IEC 17025:2005 compliance certification for Ohio Semitronics, Inc. products.

A-7003-06 C of C with ANSI Z540 certificate of Calibration.

Includes: Contracted Calibration service provider to provide ANSI/NCSL Z540-1-1994 compliance certification for Ohio Semitronics, Inc. products.

rev. 10-16-24